

# Statement of Product Conformity

BM Polyco Ltd  
Crown Road  
Enfield  
EN1 1TX  
United Kingdom

Certified to ISO 9001 & ISO 13485

declare that the new Medical Device and PPE described hereafter



## FINITE P INDIGO AF

Non-sterile, powder-free, synthetic nitrile rubber  
disposable examination glove

is in conformity with the provisions of Council Directive 93/42/EEC and with National Standards transposing harmonised standards EN455-1, EN455-2, EN455-3, EN455-4 and is self-certified as a Class 1 non-sterile medical device

|                      |            |
|----------------------|------------|
| Freedom from holes:  | AQL 1.5    |
| Physical properties: | 6N minimum |

---

is in conformity with the provisions of Council Directive 89/686/EEC and where such is the case, with National Standards transposing harmonised standards EN374-1:2003, EN374-2:2003, EN374-3:2003 and EN420:2003.

is identical to the PPE which is the subject of EC certificate of conformity No. 3766 issued by SATRA Technology Centre, Wyndham Way, Telford Way, Kettering, Northamptonshire, NN16 8SD, United Kingdom (Notified Body No. 0321).

conforms to the provisions of Council Directive 89/686/EEC Article 11B under the supervision of the Notified Body, LRQA Ltd, Hiramford, Middlemarch Office Village, Siskin Drive, Coventry, CV3 4FJ (Notified Body No. 0088).

**MICRO ORGANISM  
HAZARDS  
EN374-2**



**CHEMICAL  
HAZARDS  
EN374-3**



CE 0088

Complex Design



### User Information

This product does not contain natural rubber latex.  
This product does not contain chemical accelerators.

### Glove Care

Storage: Store below 40°C in dry conditions away from direct sunlight and heat.

You are advised to retain this packaging for reference

### Range Available: sizes and codes

| Sizes | 5.5        | 6.5        | 7.5        | 8.5        | 9.5        |
|-------|------------|------------|------------|------------|------------|
| Code  | MFNP100/00 | MFNP100/01 | MFNP100/02 | MFNP100/03 | MFNP100/04 |

### Food Contact



Finite P Indigo AF have been tested in accordance with EN1186. They are suitable for use with all food categories in situations of short term, repeat contact.